



POINT OF USE MIXED BED D.I. FILTRATION SYSTEMS



SINGLE UNITS



TWIN UNITS

MADE WITH



OR



PURE POLYPROPYLENE **FILLED POLYPROPYLENE**

STANDARD FEATURES

- ASSEMBLED IN CLEAN ROOM CONDITIONS
- EACH ASSEMBLY PACKED IN A SEALED PLASTIC BAG
- WALL MOUNTING POWDER COATED BRACKET AND DISASSEMBLY WRENCH
- THREADED IN/OUT PORTS
- LABORATORY GRADE VIRGIN 40C:60A RATIO
- 18 MEG PRODUCT AT RATED FLOWS ON MUNICIPAL OR R.O. WATER SOURCES

<u>MODEL</u>	<u>SUMP SIZE</u>	<u>SINGLE/TWIN (2)</u>	<u>GPM/(LPM) (1)</u> <u>FLOW RATE</u>	<u>CAPACITY-KGR</u>	<u>IPS IN/OUT</u>
DI-1	2X10	SINGLE	0.1/ 0.3 (0.4/1.2)	0.3	1/4
DI-2	2X10	TWIN	0.1/0.3 (0.4/1.2)	0.6	1/4
DI-3	2X20	SINGLE	0.3/0.9 (1.2/4.0)	0.7	1/2
DI-4	2X20	TWIN	0.3/0.9 (1.2/4.0)	1.4	1/2
DI-5	4X10	SINGLE	0.3/0.9 (1.2/4.0)	0.8	1
DI-6	4X10	TWIN	0.3/0.9 (1.2/4.0)	1.6	1
DI-7	4X20	SINGLE	0.7/2.1 (2.8/8.4)	1.9	1
DI-8	4X20	TWIN	0.7/2.1 (2.8/8.4)	3.8	1

NOTES: (1) FIRST NUMBER IS ON RAW MUNICIPAL WATER; SECOND NUMBER IS ON R.O. WATER
(2) UNITS ARE IN SERIES SO SINGLE FLOWS APPLY

OPTIONS

1. SPECIFY VIRGIN UNFILLED OR TALC FILLED POLYPROPYLENE HOUSINGS
2. 2 MEG GREEN GOOD/RED BAD LIGHT BETWEEN TWIN FILTERS TO WARN OF PENDING EXHAUSTION OF FIRST FILTER
3. WALL MOUNT DIGITAL OUTLET RESISTIVITY METER WITH LIGHT ALARM
4. RE-CIRCULATION PUMP TO MAINTAIN QUALITY